



校准证书

CALIBRATION CERTIFICATE

校准证书编号: YFJC/01-第 2018-003812 号

Calibrated certificate series No.

委 托 者 常州仁成金属制品有限公司
 Customer

委 托 者 地 址 /
 Address of customer

器 具 名 称 游标卡尺
 Name of instrument

制 造 厂 /
 Manufacturer

型 号 / 规 格 (0~200)mm
 Model/Specification

器 具 编 号 CZRC-001
 NO. of instrument



证书批准人 邵明
 Approved by

核 验 员 沈解彦
 Checked by

校 准 员 施根梯
 Calibrated by

校准日期 2018 年 10 月 18 日
 Date for calibrated Year Month Day

本证书提供的测量结果仅对本次被校对象有效。 The result of the certificate is effect only on this calibrated instrument (s).
 未经许可，不得部分复制本证书。 This certificate can not be partly if not allowed.

地址: 上海市闵行区北翟路 1551 号 邮政编码: 201100 电话: 62388221 传真: 52191183
 Address: Shanghai Minhang district Beidi road No.1551 Post Code: 201100 Tel: 62388221 Fax: 52191183





上海英菲计量设备检测有限公司

SHANGHAIYINGFEI MEASUREMENT EQUIPMENT EXAMINATION CO.,LTD.

校准证书编号: YFJC/01-第 2018-003812 号

Calibrated certificate series No.

上海市计量校准实验室认可证书号: **SJ052**

The number of the certificate accredited by Shanghai Laboratory of Measurement and Calibration is No.

本次校准所依据的技术规范(代号、名称):

Reference documents for the calibration(code、name)

JJG30-2012 通用卡尺检定规程

YFJC/BQDD-01-01 游标量具示值误差测量结果不确定度评定

本次校准所使用的主要标准器具:

Main measurement standards used in this calibration

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度 Measuring range/accuracy (10~291.8)mm/ 5 等
量块/12 块组	37002	L3-12120520/2019.05.19	

出证机构:上海市计量测试技术研究院

Issuing body: Shanghai metrological testing technology research institute

以上计量标准器具的量值溯源至国家基准。

Quantity values of above measurement standards used in this calibration are traced to those of the national primary standards in the P.R.China.

校准地点及环境条件:

Location and environmental condition for the calibration

地点: 现场

Location

温度: 22℃;

Ambient temperature

湿度: 58%RH;

Relative humidity

其它: /

Others

本次校准结果的扩展不确定度:

Expanded uncertainty

$U=0.01\text{mm}$ ($k=2$)

被校器具限制使用范围及条件:

The liminary scope and condition

/

校准证书续页专用

Continued page of calibration certificate



第 2 页 共 3 页

Page of total pages



上海英菲计量设备检测有限公司

SHANGHAIYINGFEI MEASUREMENTEQUIPMENT EXAMINATIONCO.,LTD.

校准证书编号:YFJC/01-第 2018-003812 号

Calibrated certificate series No.

校准结果/说明:

Results of calibration and additional explanation

受检点 (mm)	示值误差 (mm)
51.2	0.00
121.5	0.02
191.8	0.02

校准内容结束

经校准符合检定规程要求

建议:下次计量校准日期为 2019 年 10 月 17 日

校准证书续页专用

Continued page of calibration certificate

第 3 页 共 3 页

Page of total pages

